

产品合格证
CERTIFICATE OF QUALITY

产品名称: KN95 学生口罩 (非医用)
Product name: KN95 Particular Respirator (Kid Mask) (Non Medical)

产品规格: 110*90mm
Product Dimensions

产品数量: 10pieces
Quantity

质检员:
inspector

生产批号: YM200509
Lot No

产品标准: GB2626-2006、EN149: 2001 + A1: 2009
Executive standard

生产日期: 2020-05-09
Production Date

保质期限: 3years
Expiry Date

产品成分: 熔喷布20%、无纺布80%
Product Composition: Melt-blown nonwoven fabric 20%
spunbond non-woven fabric 80%

生产厂家: 佛山一美护理用品有限公司
Manufacturer: Foshan YiMei Hygiene Products Co., Ltd.

地址: 佛山市南海区丹灶镇华南五金产业基地明沙北路30号
Address: 30, MingShabeilu, Danzao Village Nanhai District,
Foshan, Guangdong Province, China

本产品经检验符合规定的质量标准要求, 准予出厂。
This product was tested in accordance with standard factory, granted.



Certificate of Compliance



No. OP200330E.FYH00042

Technical Construction File no. TPPEMJ2003273749

Certificate's Holder: Foshan YiMei Hygiene Products Co., Ltd.
30, MingShabeilu, Danzao Village, Nanhai District, Foshan, Guangdong Province, China. 528216

Certification ECM Mark:



Product: KN 95 Particular Respirator (Not Sterile)
Model(s): YM-KN9501, YM-KN9502, YM-KN9503, YM-KN9504, YM-PM01, YM-PM02, YM-PM03, YM-PM04.

Verification to: Standard: EN 149: 2001+A1:2009

related to CE Directive(s): R 2016/425 (Personal Protective Equipment)

Remark: This document has been issued on a voluntary basis and upon request of the manufacturer. It is our opinion that the technical documentation received from the manufacturer is satisfactory for the requirements of the ECM Certification Mark. The conformity mark above can be affixed on the products accordingly to the ECM regulation about its release and its use.

Additional information and clarification about the Marking:



The manufacturer is responsible for the CE Marking process, and if necessary, must refer to a Notified Body. This document has been issued on the basis of the regulation on ECM Voluntary Mark for the certification of products. RG01_ECM rev.3 available at: www.entecerma.it

Issuance date: 30 March 2020

Expiry date: 29 March 2025

Reviewer
Technical expert
Amanda Payne

Approver
ECM Service Director
Luca Berjoni

Ente Certificazione Macchine Srl

Via Ca' Bella, 243 – Loc. Castello di Serravalle – 40053 Valsamoggia (BO) - ITALY
☎ +39 051 6705141 📠 +39 051 6705156 ✉ info@entecerma.it 🌐 www.entecerma.it

检验检测报告

(电子版)



No:200052055

共3页 第1页

委托单位	广州新致医疗器械有限公司 地址: 广州市白云区云城街齐富路34号713房 (自主申报)		
样品信息	KN95学生口罩 30个		
检验性质	委托检测	样品受理/测试开始日期	2020-04-06
		报告签发日期	2020-04-11
判定依据	GB 2626-2006 《呼吸防护用品 自吸过滤式防颗粒物呼吸器》		
综合检验结论	-		
检验检测结果	检验检测项目	判定依据	判定
	NaCl颗粒物过滤效率	GB 2626-2006	符合
	吸气阻力	GB 2626-2006	符合
	呼气阻力	GB 2626-2006	符合
	可燃性	GB 2626-2006	符合
	头带	GB 2626-2006	符合
备注	<p>本报告中检验检测项目均在相应标准规定的环境下进行(有注明的除外)。 复印件, 副本未重新加盖报告书确认章无效。 本报告检验检测地址为广州市番禺区珠江路1号。</p>		

签发: 刘国 工程师



样品图片

(电子版)

No: 200052055

共3页 第2页



检验检测报告附页 (电子版)

No: 200052055

共3页 第3页

检验检测项目 (计量单位) (样品识别)	测试方法	标准值及允差	检验检测结果	判定	备注
●NaCl颗粒物过滤效率	GB 2626-2006 6.3 空气流量:85L/min 气溶胶颗粒:NaCl 气溶胶浓度:15mg/m ³ 温度:23.1℃ 相对湿度:36.2%	过滤效率(%): ≥95.0 (KN95)	过滤效率(%): 未处理样品: 1# 99.842 2# 99.911 3# 99.875 4# 99.798 5# 99.814 6# 99.762 7# 99.789 8# 99.815 9# 99.882 10# 99.795 温湿度预处理后样品: 1# 99.711 2# 99.714 3# 99.702 4# 99.684 5# 99.754	符合	
●吸气阻力(Pa)	GB 2626-2006 6.5 头模:中号	≤350	未处理样品: 1# 63.8 2# 67.2 预处理样品: 1# 61.1 2# 65.9	符合	
●呼气阻力(Pa)	GB 2626-2006 6.6 头模:中号	≤250	未处理样品: 1# 72.7 2# 75.0 预处理样品: 1# 73.6 2# 75.1	符合	
●可燃性(s)	GB 2626-2006 6.15	续燃时间 ≤5	续燃时间 未处理样品: 1# 0.0 2# 0.0 温湿度预处理后样品: 3# 0.0 4# 0.0	符合	
●头带	GB 2626-2006 6.11	面罩的每条头带、带扣及其他调节部件在承受10N的拉力,持续10s,不应出现滑脱或断裂。	未处理样品: 1# 符合要求 温湿度预处理后样品: 1# 符合要求	符合	
备注	(本栏空白)				



——本报告结束——